

## **“Playing around” with Field-Effect Sensors on the Basis of EIS Structures, LAPS and ISFETs**

**Michael J. Schöning** <sup>1,2,\*</sup>

<sup>1</sup> University of Applied Sciences Aachen (Division Jülich), Laboratory for Chemical Sensors and Biosensors, Ginsterweg 1, 52428 Jülich, Germany

<sup>2</sup> Institute of Thin Films and Interfaces, Research Centre Jülich, 52425 Jülich, Germany

\* Author to whom correspondence should be addressed. Email: [m.j.schoening@fz-juelich.de](mailto:m.j.schoening@fz-juelich.de)

*Received: 01 June 2004 / Accepted: 21 January 2005 / Published: 31 March 2005*

---

**Abstract:** Microfabricated semiconductor devices are becoming increasingly relevant, also for the detection of biological and chemical quantities. Especially, the “marriage” of biomolecules and silicon technology often yields successful new sensor concepts. The fabrication techniques of such silicon-based chemical sensors and biosensors, respectively, will have a distinct impact in different fields of application such as medicine, food technology, environment, chemistry and biotechnology as well as information processing. Moreover, scientists and engineers are interested in the analytical benefits of miniaturised and microfabricated sensor devices. This paper gives a survey on different types of semiconductor-based field-effect structures that have been recently developed in our laboratory.

**Keywords:** field-effect sensor, ISFET, EIS, LAPS.

---

## Introduction

The rapid development of semiconductor micro- and nano-technologies has stimulated the creation of new sensor concepts that combines both chemical and biological recognition processes with silicon chip manufacturing. Typical examples therefore, are “lab on a chip” devices,  $\mu$ TAS (micro total analysis systems), electronic tongues, etc. [1,2]. In spite of these high-sophisticated multi-parameter sensor systems, the chemical sensors or biosensors, included in the respective set-up, play the key role with regard to their analytical behaviour.

Among the variety of concepts and different types of (bio-)chemical sensors, proposed in literature, the strategy to integrate the particular chemical or biological recognition element together with a semiconductor field-effect device, is one of the most attractive approaches. In this context, capacitive EIS (electrolyte-insulator-semiconductor) sensors [3], LAPS (light-addressable potentiometric sensors) [4] and ISFETs (ion-sensitive field-effect transistors) [5], represent typical examples. These three kinds of devices are currently being the basic structural element in a new generation of chemical and biological microsensors. The main reason is that they provide a lot of potential advantages such as a small size and weight, a fast response time, the possibility of an on-chip integration of sensor arrays, a high robustness, the possibility of low-cost fabrication, etc. Moreover, their possible field of applications reaches from medicine, biotechnology, process control and environmental monitoring through food and drug industries to defence and security requirements.

This paper gives an overview on different kinds of silicon-type field-effect chemical sensor and biosensor approaches that are based on:

- thin dielectric materials in the nm-scale for ion-selective sensing (e.g., pH), prepared by pulsed laser deposition technique;
- strategies to immobilise different enzymes onto Si chips for biosensing as well as ionophores for chemical sensing;
- three-dimensionally structured porous Si as transducer material for chemical and biosensor applications;
- a “high order” hybrid FET (field-effect transistor) module for simultaneous (bio-)chemical and physical sensing;
- the development of biohybrid sensors by immobilising living cells or intact chemoreceptors.


## Detection principles

All sensors described in this work are basing on field-effect devices, either electrolyte-insulator-semiconductor (EIS) structures and light-addressable potentiometric sensors (LAPS) or ISFET- (ion-selective FET) type sensors. The set-up of these sensors is similar to that of conventional solid-state transducers such as metal-oxide-semiconductor structures.

The ISFET, EIS sensor and LAPS (see Fig. 1) are very sensitive for any kind of potential generation at or near the gate insulator/electrolyte interface, i.e. at the sensitive “gate” region. Therefore, it will be clear that each biological or chemical reaction, leading to chemical or electrical changes at this interface, can be measured by means of these devices coupled with the respective chemical or biological recognition element. Generally, the following basic mechanisms of potential generation can

be considered for this type of electrochemical sensor [6]:

- a pH or ion-concentration change,
- enzymatic reactions,
- affinity binding of molecules (antigen-antibody affinity reaction, or DNA hybridisation),
- and potential changes that are coming from living biological systems as a result of more sophisticated biochemical processes (action potential of nerve cells, dipole potentials, etc.).




**Figure 1.** Detection principles of ISFET (a), EIS sensor (b) and LAPS (c), and corresponding change of respective sensor signal ( $V_G$ : gate voltage,  $V_{DS}$ : drain-source voltage,  $I_D$ : drain current,  $V_{\sim}$ : ac voltage,  $V_{=}$ : dc voltage, WE: working electrode, RE: reference electrode).

As can be seen from Fig. 1, all these “effects” yield a change in the input characteristics of the ISFET or BioFET (biologically modified FET) or the capacitance-voltage curve of the EIS sensor or the photocurrent-voltage response of the LAPS, respectively, that are shifted along the voltage axis [7]. By measuring this voltage shift, the analyte concentration or composition to be detected can be determined.


#### EIS sensor and LAPS

For the capacitive field-effect sensors, different measuring principles have been realised, which are shown in Fig. 2: capacitance / voltage (C/V) and constant capacitance (ConCap) mode [8] as well as photocurrent / voltage (I/V) and constant photocurrent LAPS (CLAPS) mode [9]. Depending on the layer set-up used and the kind of desired application, a multitude of possible variants of the field-effect sensor has been summarised in Fig. 3. In all cases, for the capacitive field-effect sensor, the metallic “gate” electrode of a conventional MIS (metal-insulator-semiconductor) structure has been replaced by the electrochemical sensor part, consisting of the particular sensitive layer, the analyte (i.e., the test sample) and the reference electrode [10].



**Figure 2.** Measuring principles of EIS sensor and LAPS, respectively.

For all field-effect sensors, on the one hand conventional techniques of silicon planar technology (thermal oxidation, physical and chemical vapour deposition, photolithography, etching, etc.) have been used. On the other hand, these conventional techniques have been combined with novel techniques of sensor preparation such as the pulsed laser deposition (PLD) to prepare pH-sensitive thin-film dielectrics [11,12], the anodic etching process to generate porous Si as carrier matrix for enzymes or cells [13,14] and specific immobilisation strategies in order to stably fixate chemically and / or biologically sensitive materials to these microelectronic chips [15-17].



**Figure 3.** Capacitive field-effect structures realised in this work (schematically), based on EIS sensors (a) and LAPS principle (b).

The PLD process is proposed as an innovative semiconductor-compatible fabrication technique in

order to realise thin-film materials for chemical sensor applications, like  $\text{Ta}_2\text{O}_5$  and  $\text{Al}_2\text{O}_3$  as pH-sensitive gate insulators with nearly-Nernstian pH behaviour [18-20].

Porous EIS sensors and LAPS can increase the long-term stability of enzymatic field-effect sensors, in spite of the very mild immobilisation procedure by means of physical adsorption [21]; porous “spots” allow to downscale EIS sensors to sub- $\mu\text{m}$  dimensions [22-24]. Moreover, defined adhesion of cells and neurones can be studied by variation of the surface of artificially structured silicon chips [25,26].

A recently introduced dual amperometric / potentiometric FIA- (flow-injection analysis) EIS biosensor with the immobilised enzyme organophosphorus hydrolase allows the distinctive detection of organophosphate pesticides such as paraoxon, parathion, methyl parathion or diazinon, which are potent neurotoxins with structural similarities to some chemical warfare agents [27-31]. For the detection of cyanide, e.g. in metal mining and metal plating industry, a cyanide-specific biosensor exploiting immobilised cyanidase has been presented with a detection limit in the  $\mu\text{M}$  concentration range [32].

EIS sensors have not been only developed with regard to environmental but also pharmaceutical applications by using, e.g. an alliin-specific biosensor based on the enzyme alliinase: enzymatically formed ammonia can be detected by the pH-sensitive EIS structure. Cystein sulfoxides can be monitored in this way for breeding research and screening purposes of potential medical plants [33,34]. A penicillin-sensitive EIS structure with a high long-term stability was fabricated by immobilising the enzyme penicillinase with a heterobifunctional cross-linker [35-38].

Ion-selective EIS structures and LAPS have been arranged in different configurations as single sensors and sensor arrays [36,39,40]: Examples are a  $\text{K}^+$ -selective EIS sensor with a valinomycin-containing PVC membrane, a PVC-based  $\text{Li}^+/\text{K}^+$  and  $\text{Ca}^{2+}/\text{Li}^+$  multi-sensor LAPS, an anion-selective LAPS for the determination of nitrate and sulphate ions and a LAPS for  $\text{Cs}^+$ ,  $\text{Mg}^{2+}$  and  $\text{Li}^+$  detection based on photocurable membranes [41-47]. Enzyme-based LAPS for the determination of urea and butyrylcholine have been elaborated with photocurable polymeric enzyme membranes of urease and BuChE (butyrylcholinesterase), respectively [48].


#### *ISFET-type sensor*

The subject of miniaturised multi-sensor systems, i.e. the combination of several sensors to sensor arrays is attracting increasing attention. In such a  $\mu\text{TAS}$ , the detector unit often is a multi-sensor module, intended for both physical and (bio-)chemical quantities, like concentration of ions, biomolecules, temperature, flow-rate, etc. Usually, a large number of single-function sensors are combined that suffer from the drawback of different sensitive layers and transducer principles, which have to be optimised and operated at the same time. In contrast, a relatively new trend is the development of so-called “high order” sensor arrays, which imply more than one transducer principle for the same chemically sensitive layer [49].

As an alternative, in our research experiment, a hybrid sensor module that is based on an identical transducer principle has been suggested. In this sensor / actuator set-up, the same ISFET can serve as both a physical and bio- / or chemical sensor. Consequently, the amount of (bio-)chemical and physical information is higher than the number of sensors that are present in the module. The “high

order” ISFET module consists of two ISFETs (either two ISFETs or one ISFET and a second BioFET), an ion generator and a reference electrode (see Fig. 4). The multi-functionality in this “high order” ISFET module is achieved by means of sequential or simultaneous scheduling of the ISFETs in different combinations and / or different ISFET operation modes [50]. The multi-parameter system allows the detection of seven chemical / biological and physical quantities such as pH, penicillin concentration, temperature, diffusion coefficient of ions, flow direction, flow velocity and liquid level [51]. A pH-sensitive Ta<sub>2</sub>O<sub>5</sub>-gate ISFET is applied as transducer for all suggested sensors [52-54]. For the measuring principle, the ISFETs within the hybrid sensor module are operated in the constant charge mode [55-58].

A further extension of the functional possibilities of the developed “high order” hybrid ISFET module in combination with a highly sensitive and selective detection of odour concentrations aims in the realisation of a bioelectronic sensor: taking whole animals or at least complete sensory organs as a biological recognition element. For example, insects are known for their extraordinary sensory abilities. Therefore, an odour-sensitive beetle/chip biosensor has been created by coupling the responsible organ for smell, i.e. the insect antenna, directly to a FET (“whole-beetle” BioFET / “isolated antenna” BioFET). In this approach, the voltage generated in the antenna upon smelling a certain odour concentration is used to modify the drain current of the transistor. Such a beetle/chip biosensor enables the highly sensitive detection of odour concentrations down to the ppt concentration range [59-67].



**Figure 4.** Possible sensor configurations for the measurement of seven (bio-)chemical / physical parameters using the ISFET-based “high order” module.

## Summary

Different types of chemical sensors and biosensors sensitive towards various ions and analytes have been developed using the ISFET, EIS structure or LAPS as transducer. They are summarised in Table 1. Some of these sensors (e.g., pH and penicillin sensors) have been also realised using a porous EIS structure or porous LAPS.

With regard to possible practical applications, e.g. in environmental monitoring and medicine, the pH-sensitive ISFET was utilised for both pH determination in rain drops and in human urine and the glucose-sensitive ISFET was used for glucose concentration measurement in urine. The penicillin-sensitive EIS sensor has been proven for the penicillin detection in fermentation processes; the pH-sensitive EIS sensor has been tested in food technology, especially with respect to a feasible CIP (cleaning in process) procedure.

In addition, momentary work deals with new concepts of integrated miniaturised reference electrodes in silicon technology for potentiometric sensor systems. Therefore, different types of reference electrodes have been realised by means of thin-film and thick-film techniques (electron-beam evaporation / pulsed laser deposition or chlorination; screen-printing) [68-70].

**Table 1.** Summary of developed (bio-)chemical sensors: EIS, LAPS, ISFET.

(Bio-)chemical sensor	Ion / analyte	Sensitive membrane or (bio-)recognition element	Transducer
pH sensor	$H^+$ , $OH^-$	$Si_3N_4$ ; $Al_2O_3$ ; $Ta_2O_5$	(porous) EIS / LAPS; ISFET
Ion sensor	$K^+$ , $Li^+$ , $Cs^+$ , $Ca^{2+}$ , $Mg^{2+}$ , $NO_3^-$ , $SO_4^{2-}$	Polymer membrane & ionophore	EIS; LAPS; ISFET
Enzyme sensor	Glucose, urea, penicillin, alliin, pesticides, cyanide, butyrylcholine	Glucose oxidase, urease, penicillinase, alliinase, organophosphorus hydrolase, cyanidase, BuChE	ISFET (porous) EIS /LAPS
Beetle/chip sensor	Cis-3-hexen-1-ol	Insect antenna	ISFET

## Acknowledgments

The author gratefully thanks Y. Ermolenko, M. Keusgen, J.P. Klock, H. Lüth, Y. Mourzina, A. Mulchandani, A. Poghosian, P. Schroth, J.W. Schultze, S. Schütz, A. Simonis, A. Steffen, M. Thust, T. Wagner, J. Wang and T. Yoshinobu for technical support and valuable discussions and the Ministerium für Wissenschaft und Forschung des Landes Nordrhein-Westfalen for partial financial support.

## References

1. Shoji, S. *Micro total analysis systems, Sensors Update*; Eds.: Baltes, H.; Göpel, W.; Hesse, J. Wiley-VCH: Weinheim, **2000**, 3-17.
2. van den Berg, A.; Lammerink, T.S.J. Micro total analysis systems: microfluidic aspects, integration concept and applications, *Topics in Current Chemistry*, **1997**, *194*, 21.
3. Klein M. Characterization of ion-sensitive layer systems with a C(V) measurement method operating at constant capacitance, *Sens. Actuators B*, **1990**, *1*, 354.
4. Hafeman D.G.; Parce J.W.; McConnel H.M. Light-addressable potentiometric sensor for biochemical systems, *Science*, **1988**, *240*, 1182.
5. Bergveld P. Development of an ion-sensitive solid-state device for electrophysiological measurements, *IEEE Trans. Biomed. Engineering*, **1970**, *17*, 70.
6. Schöning, M.J.; Poghosian, A. Recent advances in biologically sensitive field-effect transistors (BioFETs), *Analyst*, **2002**, *127*, 1137.
7. Poghosian, A.; Yoshinobu, T.; Simonis, A.; Ecken, H.; Lüth, H.; Schöning, M.J. Penicillin detection by means of field-effect based sensors: EnFET, capacitive EIS sensor or LAPS?, *Sensors and Actuators B*, **2001**, *78*, 237.
8. Poghosian, A.; Thust, M.; Schroth, P.; Steffen, A.; Lüth, H.; Schöning, M.J. Penicillin detection by means of silicon-based field-effect structures, *Sensors and Materials*, **2002**, *13* (4), 207.
9. Yoshinobu, T.; Ecken, H.; Poghosian, A.; Simonis, A.; Iwasaki, H.; Lüth, H.; Schöning, M.J. Constant-current-mode LAPS (CLAPS) for the detection of penicillin, *Electroanalysis*, **2001**, *13* (8-9), 733.
10. Schöning, M.J.; Glück, O.; Thust, M. In *The measurement, Instrumentation and Sensors Handbook*; Ed.: Webster, J.G. CRC Press: Boca Raton, **1999** and Springer-Verlag: Heidelberg, 70/1-70/49.
11. Schöning, M.J.; Mourzina, Yu.G.; Schubert, J.; Zander, W.; Legin, A.; Vlasov, Yu.G.; Lüth, H. Can pulsed laser deposition serve as an advanced technique in fabricating chemical sensors?, *Sensors and Actuators B*, **2001**, *78*, 273.
12. Schöning, M.J.; Mourzina, Yu.G.; Schubert, J.; Zander, W.; Legin, A.; Vlasov, Yu.G.; Lüth, H. Pulsed laser deposition – an innovative technique for preparing inorganic thin films, *Electroanalysis*, **2001**, *13* (8-9), 727.
13. Thust, M.; Schöning, M.J.; Frohnhoff, S.; Arens-Fischer R.; Kordos, P.; Lüth, H. Porous silicon as a substrate material for potentiometric biochemical sensors, *Measurements Science and Technology*, **1996**, *7*, 26.
14. Kurowski, A.; Schultze, J.; Lüth, H.; Schöning, M.J. Micro- and nanopatterning of sensor chips by means of macroporous silicon, *Sensors and Actuators B*, **2002**, *83*, 123.



15. Schöning, M.J. in: *Electrochemical Microsystem Technologies, New Trends in Electrochemistry*, Eds.: Schultze, J.W.; Osaka, T.; Datta, M. Taylor & Francis: London, New York, **2002**; Vol. 2, 384-408.
16. Schöning, M.J.; Lüth, H. Novel concepts for silicon-based biosensors, *Physica Status Solidi A*, **2001**, 185 (1), 65.
17. Schöning, M.J.; Thust, M.; Kordos, P.; Lüth, H. In *Advances in Science and Technology*, Eds.: Vinenzini, P.; Dori, L. *Solid State Chemical and Biochemical Sensors*, Techna Srl., **1999**, 26, 55.
18. Schöning, M.J.; Tsarouchas, D.; Schaub, A.; Beckers, L.; Zander, W.; Schubert, J.; Kordos, P.; Lüth, H. A highly long-term stable silicon-based pH sensor using pulsed laser deposition technique, *Sensors and Actuators B*, **1996**, 35, 228.
19. Schöning, M.J.; Lüth, H. in: *Coupling of Biological and Electronic Systems*, Ed.: Hoffmann, K.-H. Springer-Verlag: Berlin, Caesarium: Bonn, **2002**, 79-92.
20. Ismail, A.B.M.; Harada, T.; Yoshinobu, T.; Iwasaki, H.; Schöning, M.J.; Lüth, H. Investigation of pulsed laser-deposited Al<sub>2</sub>O<sub>3</sub> as a high pH-sensitive layer for LAPS-based biosensing applications, *Sensors and Actuators B*, **2000**, 71, 169.
21. Schöning, M.J.; Ronkel, F.; Crott, M.; Thust, M.; Schultze, J.W.; Kordos, P.; Lüth, H. Miniaturization of potentiometric sensors using porous silicon microtechnology, *Electrochimica Acta*, **1997**, 47 (22), 3185.
22. Schöning, M.J.; Kurowski, A.; Thust, M.; Kordos, P.; Schultze, J.W.; Lüth, H. Capacitive microsensors for biochemical sensing on porous silicon technology, *Sensors and Actuators B*, **2000**, 64, 59.
23. Schöning, M.J.; Malkoc, Ü.; Thust, M.; Steffen, A.; Kordos, P.; Lüth, H. Novel electrochemical sensors with structured and porous semiconductor/insulator capacitors, *Sensors and Actuators B*, **2000**, 65, 288.
24. Thust, M.; Schöning, M.J.; Schroth, P.; Malkoc, Ü.; Dicker, C.I.; Steffen, A.; Kordos, P.; Lüth, H. Enzyme immobilisation on planar and porous silicon substrates for biosensor applications, *Journal of Molecular Catalysis B: Enzymatic*, **1999**, 7, 77.
25. Schöning, M.J.; Simonis, A.; Ruge, C.; Ecken, H.; Müller-Veggian, M.; Lüth, H. A (bio-) chemical field-effect sensor with macroporous Si as substrate material and a SiO<sub>2</sub> / LPCVD-Si<sub>3</sub>N<sub>4</sub> double layer as pH transducer, *Sensors*, **2002**, 2, 11.
26. Heiduschka, P.; Rommann, I.; Ecken, H.; Schöning, M.J.; Schuhmann, W.; Thanos, S. in: *Scaling Down in Electrochemistry: Electrochemical Micro- and Nanosystem Technology*, *Electrochimica Acta* 47, Eds.: Schultze, J.W.; Staikov, G. Pergamon Press Elsevier: Amsterdam, **2001**, 299-307.
27. Schöning, M.J.; Krause, R.; Block, K.; Musameh, M.; Mulchandani, A.; Wang, J. A dual amperometric/ potentiometric FIA-based biosensor for the distinctive detection of organophosphorus pesticides, *Sensors and Actuators B*, **2003**, 95, 291.

28. Schöning, M.J.; Arzdorf, M.; Mulchandani, P.; Chen, W.; Mulchandani, A. Towards a capacitive enzyme sensor for direct determination of organophosphorus pesticides: Fundamentals studies and aspects of development, *Sensors*, **2003**, *3*, 119.
29. Schöning, M.J.; Arzdorf, M.; Mulchandani, P.; Chen, W.; Mulchandani, A. A capacitive field-effect sensor for the direct determination of organophosphorus pesticides, *Sensors and Actuators B*, **2003**, *91*, 92.
30. Wang, J.; Krause, R.; Block, K.; Musameh, M.; Mulchandani, A.; Schöning, M.J. Flow injection amperometric detection of OP nerve agents based on organophosphorus-hydrolase biosensor detector, *Biosensors & Bioelectronics*, **2003**, *18*, 255.
31. Wang, J.; Krause, R.; Block, K.; Musameh, M.; Mulchandani, A.; Mulchandani, P.; Chen, W.; Schöning, M.J. Dual amperometric-potentiometric biosensor detection system for monitoring organophosphorus neurotoxins, *Analytica Chimica Acta*, **2002**, *469*, 197.
32. Keusgen, M.; Jünger, M.; Krest, I.; Schöning, M.J. Direct determination of cyanides by potentiometric biosensors, *Proc. Eurosensors XVII*, Guimaraes (Portugal) 21-24 September **2003**, 817.
33. Keusgen, M.; Jünger, M.; Krest, I.; Schöning, M.J. Biosensoric detection of the cystein sulfoxide alliin, *Sensors and Actuators B*, **2003**, *95*, 297.
34. Keusgen, M.; Jünger, M.; Krest, I.; Schöning, M.J. Development of a biosensor specific for cysteine sulfoxides, *Biosensors & Bioelectronics*, **2003**, *18*, 805.
35. Thust, M.; Schöning, M.J.; Vetter, J.; Kordos, P.; Lüth, H. A long-term stable penicillin-sensitive potentiometric biosensor with enzyme immobilized by heterobifunctional crosslinking, *Analytica Chimica Acta*, **1996**, *323*, 115.
36. Schöning, M.J.; Thust, M.; Müller-Veggian M.; Kordos, P.; Lüth, H. A silicon-based sensor array with capacitive EIS structures, *Sensors and Actuators B*, **1998**, *47*, 224.
37. Poghossian, A.; Thust, M.; Schöning, M.J.; Müller-Veggian, M.; Kordos, P.; Lüth, H. Cross-sensitivity of a capacitive penicillin sensor combined with a diffusion barrier, *Sensors and Actuators B*, **2000**, *68*, 260.
38. Poghossian, A.; Schöning, M.J.; Schroth, P.; Simonis, A.; Lüth, H. An ISFET-based penicillin sensor with high sensitivity, low detection limit and long lifetime, *Sensors and Actuators B*, **2001**, *76*, 519.
39. Yoshinobu, T.; Ecken, H.; Poghossian, A.; Lüth, H.; Iwasaki, H.; Schöning, M.J. Alternative sensor materials for light-addressable potentiometric sensors, *Sensors and Actuators B*, **2001**, *76*, 388.
40. Simonis, A.; Ruge, C.; Müller-Veggian, M.; Lüth, H.; Schöning, M.J. A long-term stable macroporous-type EIS structure for electrochemical sensor applications, *Sensors and Actuators B*, **2003**, *91*, 21.

41. Yoshinobu, T.; Ecken, H.; Ismail, Md.A.B.; Iwasaki, H.; Lüth, H.; Schöning, M.J. in: *Scaling Down in Electrochemistry: Electrochemical Micro- and Nanosystem Technology*, *Electrochimica Acta* **47**, Eds.: Schultze, J.W.; Staikov, G. Pergamon Press Elsevier: Amsterdam, **2001**, 259-263.
42. Yoshinobu, T.; Schöning, M.J.; Otto, R.; Furuichi, K.; Mourzina, Yu.; Ermolenko, Yu.; Iwasaki, I. Portable light-addressable potentiometric sensor (LAPS) for multisensor applications, *Sensors and Actuators B*, **2003**, *95*, 352.
43. Mourzina, Yu.; Mai, Th.; Poghossian, A.; Ermolenko, Yu.; Yoshinobu, T.; Vlasov, Yu.; Iwasaki, H.; Schöning, M.J.  $K^+$ -selective field-effect sensors as transducers for bioelectronic applications, *Electrochimica Acta*, **2003**, *48*, 3333.
44. Mourzina, Yu.G.; Ermolenko, Yu.E.; Yoshinobu, T.; Vlasov, Yu.; Iwasaki, H.; Schöning, M.J. Anion-selective light-addressable potentiometric sensors (LAPS) for the determination of nitrate and suphate ions, *Sensors and Actuators B*, **2003**, *91*, 32.
45. Ermolenko, Yu.; Yoshinobu, T.; Mourzina, Yu.; Furuichi, K.; Levichev, S.; Schöning, M.J.; Vlasov, Yu.; Iwasaki, H. The double  $K^+/Ca^{2+}$  sensor based on laser scanned silicon transducer (LSST) for multi-component analysis, *Talanta*, **2003**, *59*, 785.
46. Ermolenko, Yu.; Yoshinobu, T.; Mourzina, Yu.; Levichev, S.; Furuichi, K.; Vlasov, Yu.; Schöning, M.J.; Iwasaki, H. Photocurable membranes for ion-selective light-addressable potentiometric sensors, *Sensors and Actuators B*, **2002**, *85*, 79.
47. Ermolenko, Yu.; Yoshinobu, T.; Mourzina, Yu.; Furuichi, K.; Levichev, S.; Vlasov, Yu.; Schöning, M.J.; Iwasaki, H. Lithium sensor based on the laser scanning semiconductor transducer, *Analytica Chimica Acta*, **2002**, *459*, 1.
48. Mourzina, I.G.; Yoshinobu, T.; Ermolenko, Y.E.; Vlasov, Y.G.; Schöning, M.J.; Iwasaki H. Immobilization of urease and cholinesterase on the surface of semiconductor transducer for the development of light-addressable potentiometric sensors, *Mikrochimica Acta*, **2004**, *144*, 41.
49. Janata, J.; Josowicz, M.; Vanysek, P.; DeVaney, D.M. Chemical sensors, *Analytical Chemistry*, **1998**, *70*, 179R.
50. Poghossian, A.; Schöning, M.J. in: *Integrated Analytical Systems, Comprehensive Analytical Chemistry XXXIX*; Ed.: Alegret, S. Elsevier: Amsterdam, **2003**, 587-623.
51. Schöning, M.J.; Schultze, J.W.; Poghossian, A., Measuring seven parameters by two ISFET modules in a microcell set-up, *Int. Journal of Computational Engineering Science*, **2003**, *4* (2), 257.
52. Poghossian, A.; Lüth, H.; Schultze, J.W.; Schöning, M.J., in: *Scaling Down in Electrochemistry: Electrochemical Micro- and Nanosystem Technology*, *Electrochimica Acta* **47**, Eds.: Schultze, J.W.; Staikov, G. Pergamon Press, Elsevier: Amsterdam, **2001**, 243-249.
53. Poghossian, A.; Schultze, J.W.; Schöning, M.J., Application of a (bio-)chemical sensor (ISFET) for the detection of physical parameters in liquids, *Electrochimica Acta*, **2003**, *48*, 3289.

54. Poghosian, A.; Yoshinobu, T.; Schöning, M.J. Flow-velocity microsensors based on semiconductor field-effect structures, *Sensors*, **2003**, *3*, 202.
55. Poghosian, A.; Schultze, J.W.; Schöning, M.J. Multi-parameter detection of (bio-)chemical and physical using an identical transducer principle, *Sensors and Actuators B*, **2003**, *91*, 83.
56. Schöning, M.J.; Poghosian, A.; Schultze, J.; Lüth, H. Field-effect based multifunctional hybrid sensor module for the determination of both (bio-)chemical and physical parameters, *Proceedings of SPIE Reprint*, **2002**, *4576*, 149.
57. Poghosian, A.; Berndsen, L.; Lüth, H.; Schöning, M.J. Novel concepts for flow-rate and flow-direction determination by means of pH-sensitive ISFETs, *Proceedings of SPIE Reprint*, **2001**, *4560*, 19.
58. Poghosian, A.; Berndsen, L.; Schöning, M.J. Chemical sensor as physical sensor: ISFET-based flow-velocity, flow-direction and diffusion-coefficient sensor, *Sensors and Actuators B*, **2003**, *95*, 384.
59. Schöning, M.J.; Schroth, P.; Schütz, S. The use of insect chemoreceptors for the assembly of biosensors based on semiconductor field-effect sensors, *Electroanalysis, Microscale Systems*, **2000**, *12* (9), 645.
60. Schroth, P.; Lüth, H.; Hummel, H.E.; Schütz, S.; Schöning, M.J. in: *Scaling Down in Electrochemistry: Electrochemical Micro- and Nanosystem Technology*, *Electrochimica Acta* **47**, Eds.: Schultze, J.W.; Staikov, G. Pergamon Press, Elsevier: Amsterdam, **2001**, 293.
61. Schroth, P.; Schöning, M.J.; Lüth, H.; Weißbecker, B.; Hummel, H.E.; Schütz, S. Extending the capabilities of an antenna/chip biosensor by employing various insect species, *Sensors and Actuators B*, **2001**, *78*, 1.
62. Schöning, M.J.; Schroth, P.; Lüth, H.; Hummel, H.E.; Schütz, S. Insect chemoreceptors coupled to silicon transistors as innovative biosensors, *Proceedings of SPIE Reprint*, **2001**, *4205*, 152.
63. Schütz, S.; Schöning, M.J.; Schroth, P.; Weißbecker, B.; Kordos, P.; Lüth, H.; Hummel, H.E. An insect-based BioFET as a bioelectronic nose, *Sensors and Actuators B*, **2000**, *65*, 291.
64. Schroth, P.; Schöning, M.J.; Schütz, S.; Malkoc, Ü.; Steffen, A.; Marso, M.; Hummel, H.E.; Kordos, P.; Lüth, H. Coupling of insect antennae to field-effect transistors for biochemical sensing, *Electrochimica Acta*, **1999**, *44*, 3821.
65. Schroth, P.; Schöning, M.J.; Kordos, P.; Lüth, H.; Schütz, S.; Weißbecker, B.; Hummel, H.E. Insect-based BioFETs with improved signal characteristics, *Biosensors & Bioelectronics*, **1990**, *14*, 303.
66. Schöning, M.J.; Schütz, S.; Schroth, P.; Weißbecker, B.; Steffen, A.; Kordos, P.; Lüth, H.; Hummel, H.E. A BioFET on the basis of intact insect antennae, *Sensors and Actuators B*, **1998**, *47*, 234.

67. Schütz, S.; Weißbecker, B.; Hummel, H.E.; Schöning, M.J.; Riemer, A.; Kordos, P.; Lüth, H. Field effect transistor - insect antenna junction, *Naturwissenschaften*, **1997**, *84*, 86.
68. Simonis, A.; Krings, T.; Lüth, H.; Wang, J.; Schöning, M.J. A „hybrid“ thin-film pH sensor with integrated thick-film reference, *Sensors*, **2001**, *1*, 183.
69. Schöning, M.J.; Simonis, A.; Krings, T.; Lüth, H.; Wang, J. Evaluation of a chip-based thin-film / thick-film sensor hybrid for (bio-)chemical analysis, *Electroanalysis*, **2002**, *14 (13)*, 955.
70. Simonis, A.; Lüth, H.; Wang, J.; Schöning, M.J. Strategies of miniaturised reference electrodes integrated in a silicon-based „one chip“ pH sensor, *Sensors*, **2003**, *3*, 202.