



Heinz Maier-Leibnitz Zentrum
Neutronen für Forschung, Industrie und Medizin

New sample environment options at the neutron diffractometer “BioDiff”

Cryostream - closed cycle cryostat - high pressure cell for powder diffraction

March 22nd 2013 | Tobias E. Schrader

Outline of this talk


- Advantages and draw backs of neutron diffraction as compared to x-ray diffraction
- Short introduction to the instrument BioDiff
- First published results on a β -lactamase crystal
- Sample environment options at the instrument BioDiff
- Summary

Advantages / disadvantages of structure determination with neutrons I

Comparison of form factors (X-ray) and scattering lengths (neutrons):

Nucleus	atomic number	scattering length [10 ⁻¹² cm]
¹ H	1	-0.378
² H	1	0.667
¹² C	6	0.665
¹⁵ N	7	0.921
¹⁶ O	8	0.581

σ_{coh} of ¹H is $1.8 \times 10^{-28} \text{ m}^2$ but
 σ_{incoh} of ¹H is $80.2 \times 10^{-28} \text{ m}^2$
Large background from hydrogen atoms!



diameters correspond to:
form factor / scattering length
(scaled for C-atom)

Advantages of structure determination with neutrons II

hydrogen atoms can be resolved even at a resolution of $d_{\min} \approx 2\text{\AA}$ (for ${}^2\text{H}$)


Therefore one can determine:

- protonation states of amino acid side chains
- deuterium exchange as a measure of flexibility and accessibility (discrimination between **H** / **D**)
- solvent structure including hydrogen atoms


Radiation damage (hydrogen abstraction, disulfide bond cleavage) observed when measuring an x-ray crystallography data set can be avoided.

Especially, the reduction of metals in metalo-proteins can be avoided.


Schematic overview over BioDiff: A neutron protein diffractometer: collaboration between JCNS and FRMII




BioDiff, the corresponding view in reality:



Side view...



Size considerations of protein crystals



deuterated
mother
liquor

sample
crystal

size:

x-ray-crystallography:

ca. $10 \mu\text{m} \times 10 \mu\text{m} \times 10 \mu\text{m}$

typically cryoprotectants needed to facilitate measurements at low (80 K) temperatures

neutron protein crystallography:


The desirable size should be around

1mmx1mmx1mm (depending on the protein/space group)

It is desirable to exchange as many hydrogen atoms by deuterium as possible to reduce the incoherent scattering background caused by hydrogen atoms in the protein.

Outer diameter of the glass tube: 2 mm


β -lactamase: hydrolyses β -lactam antibiotics



The catalytic cycle of a class A β -lactamase illustrated for a cephalosporin substrate (inside box) and the mode of inhibition by BZB (outside box). The general base employed is not necessarily the same for acylation and deacylation. The overall reaction pathway for β -lactam hydrolysis of a cephalosporin-like substrate by the class A β -lactamase enzymes.

Figures from: Tomanicek et al., doi: 10.1074/jbc.M112.436238

Comparing the first user data set on β -lactamase with the structure recorded on LADI III @ ILL



The active center of β -lactamase
as seen by the BioDiff data set

Data collection at room temperature.





Fig. 2. Protonation states of the active site residues adjacent to Glu166 (a) and Lys234 (b) identified by σ_A -weighted $2F_o - F_c$ positive nuclear density maps colored in blue and contoured at positive 1.2σ . Nuclear density corresponding to the orientation of the catalytic water molecule (wat1) is also shown.

Fig. 2b of S. J. Tomanicek, et al. and Leighton
Coates: FEBS Letters 585 (2011) 364

Catalytic proton network of the Toho-1 β -lactamase complexed with BZB

BZB=benzothiophene-2-boronic acid (BZB)




Figures from: Tomanicek et al., doi: 10.1074/jbc.M112.436238

Sample Environment options besides room temperature operation


Oxford
Cryosystems
cryostream
nozzle

Cryostream with or without κ - goometry



Nitrogen gas stream temperature range: 80 K to 500 K

Improved κ -geometry sample holder




ϕ_{AC} and κ_{AC} moved manually, keeping the crystal in the cryostream

Closed Cycle Cryostat for temperatures down to 4 K


CCD-auxiliary detector can be used to align the crystal in the beam

sample crystal glued with GE-varnish




Cryostat temperature range: 4 K to 300 K (in vacuum)

BioDiff as a powder diffractometer



high pressure cell



NAG powder sample at 2.7 Å

Summary

- The instrument BioDiff is a monochromatic neutron diffractometer optimised for crystals with large unit cells.
- Besides room temperature BioDiff can offer a Cryostream which allows for temperatures between 90 K and 500 K.
- A closed cycle cryostat is available for temperatures down to 4 K.
- BioDiff can also be used as a powder diffractometer for crystal powders with large unit cell dimensions.

Thanks to... ... the BioDiff-Team:

- Philipp Jüttner
- Andreas Ostermann
- Reinhard Schätzler
- Bernhard Laatsch
- Frank Suxdorf
- Manfred Bednarek
- Matthias Drochner
- Harald Kleines
- Kevin Körrentz
- Karl-Heinz Mertens
- Michael Monkenbusch
- Michael Wagener
- Heinrich Pohl
- Vladimir Ossovyyi
- Andreas Nebel
- Simon Staringer
- Winfried Petry
- Severin Denk
- Dieter Richter

and you for your attention!


The End

Comparison between a data set recorded at BioDiff and LADI III

β-lactamase	LADI III @ ILL ¹	BioDiff @ FRM II
used beam time	?	8.3 days
Unit-cell parameters (Å)	a=72.92, b=72.92, c=98.53; $\alpha=\beta=90^\circ$, $\gamma=120^\circ$	a=73.26, b=73.26, c=98.66; $\alpha=\beta=90^\circ$, $\gamma=120^\circ$
Space group	P3 ₂ 1	P3 ₂ 1
No. of unique reflections	14 991	18 914
Resolution range (Å)	63.15-2.10 (2.21-2.10)	50-2.0 (2.07-2.0)
Multiplicity	7.2 (6.2)	2.0 (1.5)
Mean ((I)/sd(I))	10.1(8.2)	6.5(3.2)
Data completeness (%)	83.7(59.4)	93.6(89.6)

Crystallographic refinement	LADI III @ ILL	BioDiff @ FRM II ²
R _{factor} (%)	22.5	23.3
R _{free} (%)	25.9	25.2

β -lactamase: binding pocket without substrate



water was not included in the refinement






Fig. 3. Summary of the hydrogen-bonding network (shown with dashed lines) and bond distances (\AA) in the active site region of the pd-Toho-1 R274N/R276N.

Switching between imageplate and CCD detector




Neutrons



Neutrons

The CCD-Detector can be used to align the crystal in the neutron beam.

Instrument Characterization: The two detectors




Very little x-ray structural studies give access to hydrogen positions

X-ray:


hydrogens visible at resolution
of $d_{\text{mir}} \leq 1\text{\AA}$

neutrons:

hydrogens visible at resolution
of $d_{\text{min}} \approx 2\text{\AA}$ (for ${}^2\text{H}$)




HYDRATION STRUCTURE:




TRIANGULAR



ELLIPSOIDAL (#10)




SPHERICAL (1)




- All types of shapes can be found on the protein surface
- Triangular shapes are only found in direct contact with the protein surface
- Shape depends on the environment

PROTONATION STATES OF AMINO ACIDS:



- Fo-Fc OMIT-map; +3.0 σ
 - Fo-Fc OMIT-map; -3.0 σ
 - 2Fo-Fc map; +1.5 σ
- OMIT: H, D



EXAMPLE HIS-97:

→ C_{ε1}-H group shows acidic behavior

Niimura N, Chatake T, Ostermann A, Kurihara K, Tanaka T. (2003) Z. Kristallogr. 218:96

Other neutron protein diffractometers in the world:

Table 2

Neutron diffractometers available for protein crystallography

Diffractometer	Method	Flux at sample position (n/(cm ² s))	Typical crystal size (mm ³)	Cell parameters (Å)	Necessary days to collect data (days)	Resolution (Å)	Power target (spallation source)	Power (reactor)
BIX-3 (JAEA)	Monochromatized 2.9 Å	2.9×10^6	>1	<100	>20	>1.4		20 MW
BIX-4 (JAEA)	Monochromatized 2.6 Å	4.5×10^6	>1	<100	>15	>1.4		20 MW
iBIX ^a (J-PARC)	Pulsed white (TOF) 25 Hz (0.7–3.8 Å) divergence	2.1×10^8 at 0.20° (0.7–3.8 Å)	>0.1	<150	>3	>1.2	1 MW Hg target	
LADI-III (ILL)	Continuous source quasi-Laue (multilayer band-pass filters with $\delta\lambda/\lambda$ from 5 to 25%)	3×10^7 using $\delta\lambda/\lambda = 20\%$ centered at $\lambda = 3.5$ Å	>0.1 (if perdeuterated) >0.5 otherwise	<160	>3	>1.4		58 MW
PCS (LANSCE) [46,47]	Pulsed white (TOF Laue) 20 Hz (0.6–6 Å) divergence	9.7×10^6 at 0.12° (0.6–6 Å)	>0.3	<180	>4	>1	0.1 MW W target	
MaNDI ^a (SNS)	Pulsed white (TOF Laue) 60 Hz, adjustable wavelength range with $\Delta\lambda \sim 2.7$ Å divergence	1.2×10^7 at 0.23° and 6.9×10^7 at 0.56° divergence	>0.1	<150	>1	>1.5	2 MW Hg target	

^a If under construction.

Blakeley et al. Current Opinion in Structural Biology 18 (2008) 593-600

Comparison with the BioDiff instrument:

	monochro- mator	λ	$\Delta\lambda/\lambda$	hor. diverge- nce	vert. diverge- nce	flux n / cm² s
FRM II BioDiff	PG002	2.4Å	2.9%	0.82°	0.7°	$9.0 \cdot 10^6$
		4.0Å	2.5%	1.2°	0.7°	$7.6 \cdot 10^6$
	Ge311	2.4Å	1.4%	0.83°	0.7°	$2.6 \cdot 10^6$
		2.7Å	1.2%	0.95°	0.7°	$2.3 \cdot 10^6$
		3.0Å	0.9%	1.0°	0.7°	$1.7 \cdot 10^6$
	Si111	2.6Å	2.1%	0.8°	1.75°	$4.5 \cdot 10^6$
LADI(III)		3.2Å	≈20%			$3.0 \cdot 10^7$