



Measurements of hydroperoxy radicals (HO_2) at atmospheric concentrations using bromide chemical ionization mass spectrometry

Sascha R. Albrecht¹, Anna Novelli¹, Andreas Hofzumahaus¹, Sungah Kang¹, Yare Baker¹, Thomas Mentel¹, Andreas Wahner¹, and Hendrik Fuchs¹

¹Forschungszentrum Jülich GmbH, Institute of Energy and Climate Research, Troposphere (IEK-8), 52428 Jülich, Germany

Correspondence: Sascha Albrecht (s.albrecht@fz-juelich.de)

Abstract.

Hydroxyl and hydroperoxy radicals are key species for the understanding of atmospheric oxidation processes. Their measurement is challenging due to their high reactivity, therefore very sensitive detection methods are needed. Within this study, the measurement of hydroperoxy radicals (HO_2) using chemical ionization combined with an high resolution time of flight mass spectrometer (Aerodyne Research Inc.) employing bromide as primary ion is presented. The 1σ limit of detection of 4.5×10^7 molecules cm^{-3} for a 60 s measurement is below typical HO_2 concentrations found in the atmosphere. The detection sensitivity of the instrument is affected by the presence of water vapor. Therefore, a water vapor dependent calibration factor that decreases approximately by a factor of 2 if the water vapor mixing ratio increases from 0.1 to 1.0 % needs to be applied. An instrumental background most likely generated by the ion source that is equivalent to a HO_2 concentration of $1.5 \pm 0.2 \times 10^8$ molecules cm^{-3} is subtracted to derive atmospheric HO_2 concentrations. This background can be determined by overflowing the inlet with zero air. Several experiments were performed in the atmospheric simulation chamber SAPHIR at the Forschungszentrum Jülich to test the instrument performance by comparison to the well-established laser-induced fluorescence (LIF) technique for measurements of HO_2 . A high linear correlation coefficient of $R^2 = 0.87$ is achieved. The slope of the linear regression of 1.07 demonstrates the good absolute agreement of both measurements. Chemical conditions during experiments allowed testing the instrument's behavior in the presence of atmospheric concentrations of H_2O , NO_x and O_3 . No significant interferences from these species were observed. All these facts are demonstrating a reliable measurement of HO_2 by the chemical ionization mass spectrometer presented.

Copyright statement.

1 Introduction

Understanding of the oxidation processes in the atmosphere requires sensitive measurements of the radical species involved. Hydroxyl radicals (OH) are the most important oxidative species and are highly reactive to most of the inorganic and organic



pollutants in the atmosphere. Primary sources of OH radicals are mainly ozone photolysis and in polluted environments also nitrous acid (HONO) photolysis can be of importance. Organic pollutants are oxidized by OH to produce organic peroxy radical species (RO₂) and also hydroperoxy radicals (HO₂). OH and HO₂ radicals are closely inter-connected by a radical chain reaction, in which OH is reformed by the reaction of HO₂ with nitric oxide (NO):



As the atmospheric lifetime of HO₂ radicals is typically up to a factor 10 longer than that of OH radicals, HO₂ can be regarded as an important chemical reservoir for hydroxyl radical (OH). Atmospheric NO concentrations are often sufficiently high to maintain an efficient OH production by the reaction of HO₂ with NO, so that R1 provides a large portion of the total OH production. Measurements of both species are needed to analyze the OH radicals budget.

- 10 The majority of the techniques currently applied to measure atmospheric concentrations of HO₂ radicals use chemical conversion, which is an indirect measurement. In chemical amplifying systems, a radical reaction cycle between OH and HO₂ is established by adding two reactants. The concentration of the product species is therefore amplified compared to the small, initial HO₂ concentration in the sampled air.

- PERoxy RadiCal Amplification (PERCA) instruments make use of NO and CO for the conversion of HO₂ to OH and OH
15 to HO₂, respectively. One NO₂ molecule is produced in each reaction cycle so that the initially small HO₂ concentration is amplified as NO₂, which is then detected by a luminol detector, fluorescence or absorption methods. Because RO₂ is also converted to HO₂ in the reaction with NO, these instruments measure the sum of RO₂ and HO₂. Typically an amplification of roughly a factor of 100 is achieved to produce a measurable amount of NO₂ (Cantrell et al., 1984; Hastie et al., 1991; Clemitshaw et al., 1997; Burkert et al., 2001; Sadanaga et al., 2004; Mihele and Hastie, 2000; Green et al., 2006; Andrés-
20 Hernández et al., 2010).

- Alternatively to CO, SO₂ can be used in the chemical amplifier system (Reiner et al., 1997; Hanke et al., 2002; Edwards et al., 2003; Hornbrook et al., 2011). The high sensitivity of CIMS measurement using NO₃⁻ as primary ion allows to detect H₂SO₄ produced in the reaction of SO₂ with OH. Amplification factors of approximately 10 are sufficient in this case. Like in the PERCA instrument, RO₂ is also converted to HO₂ in the reaction with NO in these instruments. However, Hornbrook et al.
25 (2011) developed a method to distinguish between HO₂ and RO₂ by operating the instrument at different chemical conditions (varying NO, SO₂ and O₂ concentrations), thereby changing the relative sensitivities for HO₂ and RO₂.

- Laser-induced fluorescence (LIF) is a sensitive technique for OH radical measurements and it is used for the indirect detection of HO₂ by its conversion into OH after reaction with NO. The concurrent conversion of some specific RO₂ radicals can contribute to the HO₂ signal (Fuchs et al., 2011; Whalley et al., 2013; Lew et al., 2018). This can be minimized by reducing
30 the NO concentration added to the sampled air for the conversion of HO₂ to OH, but on the cost of a reduced sensitivity. A comparison of three LIF instruments in 2010 before the RO₂ interference was discovered showed significant differences in measured HO₂ concentration in experiments in the SAPHIR chamber (Fuchs et al., 2010). This could have been partly due to interferences from RO₂, but measurements also differed depending on the water vapor concentration.



Several drawbacks are connected with existing HO₂ detection methods. The PERCA systems exhibit a strong water vapor dependence of the amplification factor. In addition, chemical conversion of HO₂ by the reaction with NO used in all instruments can lead to the concurrent conversion of RO₂.

Previous work by Veres et al. (2015) showed that HO₂ radicals can be detected with a CIMS instrument using iodide as primary ion. Sanchez et al. (2016) demonstrated for the first time that this approach can also be used with Br⁻. The HO₂ radicals are directly measured by a mass spectrometer as an ion cluster formed with bromide ions. In this study, the direct measurement of atmospheric concentrations of HO₂ radicals using Br-CIMS is presented. A detailed characterization of the instrument has been performed. Further, the inter-comparison with an LIF based HO₂ measurement is used to identify potential interferences.

2 Methods

2.1 Chemical ionization mass spectrometry technique

The instrument used for the detection of the Br⁻ · HO₂ cluster consists of a custom-built ion flow tube (Fig. 1) that is mounted upstream of a commercial, high resolution time-of-flight mass spectrometer (TOF-MS, Aerodyne Res.). For the detection of reactive HO₂ radicals, losses in inlets can play a significant role. As radical species are easily lost by contact on walls, the inlet of the instrument is designed to sample air directly into the ion flow tube without additional inlet lines. The TOF-MS is equipped with an atmospheric pressure ionization (API) transfer stage providing the ion transfer from the ion flow tube to the detector. The TOF mass analyzer (Tofwerk Ag, Switzerland) has a mass resolution better than 2000.

Ambient air containing HO₂ (flow rate 3.4 slm) is sampled through a 0.7 mm skimmer nozzle and is mixed with the bromide ions in the ion flow tube shown in Fig. 1. The ion flow tube has an inner diameter of 22 mm and a length of 130 mm. The distance between the ion source and the nozzle downstream is 100 mm. The ion flow tube is kept at a constant pressure of 120 hPa using a butterfly control valve upstream of a scroll pump. Assuming that 5.4 slm of gas are passing through the ion flow tube without considering the complex fluid dynamics in the ion flow tube, the mean residence time is 4 ms. Longer versions of the ion flow tube of up to twice its size were tested, but a reduced sensitivity for HO₂ was found. Downstream of the ion flow tube, the sampled air enters a commercially available transfer stage (CI-API transfer stage, Aerodyne Research Inc.) through a nozzle with 0.5 mm diameter. The transfer stage consists of two quadrupoles and direct current transfer optics that guide the ions to the TOF analyzer.

Bromide ions easily clusters with polar species e.g. acids (Caldwell et al., 1989). This enables their detection in the gas phase including HO₂, which is a relative strong acid (the binding energy is 353 kcal mol⁻¹ Harrison (1992)). In order to produce Br⁻ ions, a gas flow of 2 slm nitrogen is mixed with 10 sccm of a 0.4 % mixture of CF₃Br in nitrogen (Air Liquide Deutschland GmbH, N₂ 99.9999 % purity). The resulting gas mixture of approximately 20 ppmv CF₃Br in nitrogen is supplied to the 370 MBq ²¹⁰Po ion source to generate bromide ions.

The isotopic pattern of bromide (approx. 1 ⁷⁹Br : 1 ⁸¹Br) provides additional information if a signal detected at a certain mass contains a cluster with bromide, because similar signals need to be contained at two masses (m/z and m/z+2). Therefore,



$\text{HO}_2 \cdot \text{Br}^-$ is detected on masses 112 and 114 with similar intensities. Both signals can be used for the data evaluation in order to improve the signal-to-noise ratio.




Figure 1. Schematic drawing of the ion flow tube, where HO_2 clusters with Br^- are formed. The ion flow-tube is mounted upstream of an Aerodyne Time-of-Flight mass spectrometer.

The data are analyzed using the following procedure. 30 mass spectra measured with a time resolution of 2 s are summed up to improve the signal-to-noise ratio (cf. Sect. 3.2). The $\text{HO}_2 \cdot \text{Br}^-$ ion cluster ion count rate (m/z 112) is normalized to the count rate of the primary ion (m/z 79). The isotopic signal at a mass-to-charge ratio of 114 and 81 are treated in the same way. The signal at both isotopic masses of the $\text{HO}_2 \cdot \text{Br}^-$ ion cluster are compared to check for possible interference from ions not containing a bromide molecule. In the following step, a water vapor dependent sensitivity is applied to convert the signal to a HO_2 concentration. Details about the water vapor dependent sensitivity are presented in Sect. 3.1. Finally a constant background is subtracted from the data. No difference in the isotopic signals was observed showing that no other molecule



(not containing bromide) is interfering. In this study, only data from one of the two isotopes (m/z 112 and 79) are discussed for simplicity.

2.2 Calibration source

For calibrating the HO₂-CIMS instrument's sensitivity the same radical source is used as for calibration of the LIF instrument that is in operation at Forschungszentrum Jülich (Fuchs et al., 2011). This is possible because the designs of the inlet nozzle and flow rates of both instruments are similar. The LIF is sampling 1.0 slm and the CIMS instrument is sampling 3.4 slm. Both flows are much smaller than the total flow through the calibration source. The calibration source provides a laminar gas stream of humidified synthetic air at a flow rate of 20 slm. The gas supply device for the calibration source allows for systematic variation of the water vapor concentration. During calibrations, the water vapor concentration is altered from 0.1 to 1.6 %, in order to determine the humidity dependence of the instrument's sensitivity. Water vapor is photolysed at 185 nm at atmospheric pressure using a penray lamp leading to the production of equal concentrations of OH and HO₂ radicals (Fuchs et al., 2011). The radical concentration that is provided by the calibration source is calculated from the UV intensity that is monitored by a photo-tube detector, the flow rate and water vapor concentration. The photo-tube signal is calibrated against ozone that is concurrently produced from oxygen photolysis by the 185 nm radiation. An absorption cell in-between the UV lamp and the photolysis region can be filled with a N₂O / N₂ mixture to vary the UV intensity, as N₂O is a strong absorber at this wavelength. If excess CO is added to the synthetic air provided to the calibration source, OH is converted to HO₂, so that the HO₂ concentration is doubled compared to the operation without CO. Typically, the calibration is performed at HO₂ concentrations between 5×10^8 and 1×10^{10} molecules cm⁻³.

2.3 HO₂ detection by laser-induced fluorescence

The LIF instrument uses two detection channels to detect OH and HO₂ simultaneously. The LIF instrument has been described in detail by Holland et al. (2003), Fuchs et al. (2011), and Tan et al. (2017).

For the HO₂ measurement, a gas stream of ambient air is expanded in to the sample cell at 4 hPa. NO is added to the sampled air for the conversion of HO₂ to OH (Reaction R1). The NO concentration is adjusted to provide a HO₂ conversion efficiency of approximately 10 % in order to minimized concurrent RO₂ conversion (Fuchs et al., 2011). The OH radicals are excited by a laser pulse at 308 nm, provided by a dye laser system. Ozone can be photolysed at 308 nm, which can lead to a small interference from ozone that is subtracted from the measured signal. For the experiments discussed here, 50 ppbv O₃ gave a signal that is equivalent to a HO₂ concentration of 3×10^6 cm⁻³. The sensitivity of the HO₂ LIF detection is water vapor dependent due to the quenching of the OH fluorescence by water. The change in the sensitivity is calculated from quenching constants. Both corrections are taken into account. The accuracy of the LIF HO₂ measurement is ± 10 % from the uncertainty of the calibration. The typical precision of measurements gives an limit of detection of 1×10^7 mol cm⁻³ (2σ) for a 80 s measurement (Tan et al., 2017).



2.4 SAPHIR

SAPHIR is an atmospheric simulation chamber at the Forschungszentrum Jülich. The chamber has been described in detail by Rohrer et al. (2005). It consists of a double-wall FEP film of cylindrical shape (length 18 m, diameter 5 m, volume 270 m³). It is equipped with a shutter system that can be opened to expose the chamber air to natural sunlight. Synthetic air used in the experiments is produced from liquid nitrogen and oxygen of highest purity (Linde, purity <99.9999 %). A combination of sensitive measurement instruments allows for studying chemical systems under well-defined, atmospheric conditions and trace gas concentrations. SAPHIR has proven to be a valuable tool for inter-comparison of different measurement techniques (Fuchs et al., 2012; Dorn et al., 2013; Fuchs et al., 2010; Apel et al., 2008), as it is ensured that all instruments can sample the same air composition.

For this study, measurements were performed during a series of experiments in the SAPHIR chamber in May and June 2017. The focus of the experiments was to study the chemistry of two classes of oxidation products of isoprene: the isoprene hydroxyhydroperoxides (ISOPOOH) and the isoprene epoxydiols (IEPOX). In addition, reference experiments without addition of VOCs, as well as experiments with isoprene were performed. These experiments were used to compare the performance of the CIMS and the LIF instrument at atmospheric HO₂ concentrations, testing various conditions, e.g. presence of ozone, NO_x species and different water concentrations.

The CIMS was mounted at the bottom of the chamber, 4 m away from the LIF instrument. The ion flow tube setup shown in Fig. 1 was directly connected to the chamber, so that the sampling nozzle was sticking into the chamber.

Data from the following instruments are used for the data evaluation and interpretation: The humidity was measured using a Picarro cavity ring-down instrument (G2401 Analyzer). NO and NO₂ were monitored by a Eco Physics chemiluminescence instrument (TR780) and ozone was detected by an UV photometer (41M, Ansyco).

3 Characterization of the HO₂-CIMS

3.1 Calibration procedure

In general, the conversion of ion count rates measured by a CIMS instrument to concentrations of the detected molecule requires regular calibrations of the sensitivity. For calibrating the HO₂ sensitivity, we utilized a radical source as described in Sect. 2.2. Figure 2 shows the measured, normalized ion count rates measured by the CIMS, when the calibration source was operated at a constant water vapor mixing ratio of 1.0 %. The HO₂ concentration was varied by changing the UV radiation intensity, which was achieved by varying the N₂O concentration in the absorption cell. A linear behavior for the normalized count rate is observed in a range of 3.0×10^8 to 1.3×10^9 HO₂ molecules cm⁻³. The slope of the linear regression gives the calibration factor of 6.8×10^{-12} cm⁻³ ncps⁻¹. The intercept of 5.1×10^{-4} ncps of the linear fit indicates a HO₂ background signal. No background correction of the CIMS signal (see below) is applied here.

Alternatively, the HO₂ provided by the calibration source can be varied by changing the water mixing ratio at constant UV intensity. The HO₂ concentration provided by the calibration source is well characterized for different water mixing ratios. This




Figure 2. Count rate of HO₂ · Br⁻ ion cluster (m/z 112) normalized to the primary ion Br⁻ (m/z 79) during sampling from the HO₂ calibration source. The HO₂ concentration provided by the source was varied by attenuating the radiation of the 185 nm radiation used to photolyse water. The water vapor mixing ratio was kept constant. The error bars are smaller than the symbols in the figure.

allows the determination of the water dependency of the CIMS. The water dependent sensitivity is defined by Eq. 1, where c is the sensitivity that depends term on the water concentration.

$$[\text{HO}_2] = c(\text{H}_2\text{O}) \frac{m/z(112)}{m/z(79)} \quad (1)$$

Figure 3 shows the sensitivity determined for each water vapor mixing ratio showing a decreasing sensitivity with increasing water vapor mixing ratio. The water dependent decrease in sensitivity is nearly linear for atmospheric relevant water mixing ratios higher than 0.1 %. Two effects contribute to the water dependence: The HO₂ ion cluster is stabilized by water during the attachment process, as water takes the access energy of the cluster rearrangement during substitution by the analyte molecule. On the other hand, the HO₂ bromide ion cluster is in a fast equilibrium with polar molecules in the gas phase. If atmospheric water vapor concentrations are present in the ion flow tube, water may substitute HO₂ in the ion cluster. The ion cluster typically has a shell of water molecules at atmospheric conditions, caused by the ions polarity (Klee et al., 2014; Derpmann et al., 2012; Albrecht et al., 2014).



As indicated in R2, an excess of water can push the reaction equilibrium in the reverse direction. Thereby, the cluster switching from HO₂ · Br⁻ to H₂O · Br⁻ causes a decrease in sensitivity.




Figure 3. Measured HO_2 sensitivity as a function of the water mixing ratio in two experiments. For the calibration, HO_2 was produced by the radical source while varying the water vapor concentration which causes a change in the HO_2 radical concentration. During the ozonolysis experiment, HO_2 was produced from the ozonolysis of 2,3-dimethyl-2-butene, which is independent of the water vapor mixing ratio. The red line shows a third order polynomial fit applied to the calibration data.

Further, the HO_2 radical itself can form a water cluster (Aloisio and Francisco, 1998; Kanno et al., 2005; Stone and Rowley, 2005). This, for example, leads to an enhancement of the HO_2 self reaction of up to a factor of two for atmospheric conditions (Stone and Rowley, 2005). However, only a fraction of the HO_2 (20 % at 297 K and 50 % humidity) is attached to a water molecule at atmospheric conditions (Kanno et al., 2005). The concentration of HO_2 water radical clusters is further reduced because of the lower pressure in the ion flow tube along with a lower partial water pressure. Therefore, compared to dry conditions an roughly 10x increased sensitivity at humid conditions is likely mainly caused by the ion water cluster.

A direct calibration for dry conditions was not possible with the radical source, because the calibration source needs water to generate HO_2 . The sensitivity of the instrument was also characterized by the production of HO_2 from the ozonolysis of 2,3 dimethyl-2-butene, that was added in a concentration of 30 ppbv to a mix of synthetic air and 200 ppbv ozone. The radical source was used as a flow-tube to overflow the inlet of the instrument with this gas mixture. 0.2 % CO was added to scavenge OH radicals produced from the ozonolysis reaction by a fast conversion of OH to HO_2 . The water mixing ratio was altered during the ozonolysis experiment from 0.0 to 0.6 %. Assuming that the HO_2 concentration from the ozonolysis is constant, the relative change in the signal gives the relative change of the instrument sensitivity. In addition, calibration measurements using the water photolysis were performed for water vapor mixing ratios higher than 0.1 %, so that the water dependence of the sensitivity determined by the two methods can be compared. As shown in Fig. 3, the instrument response is similar in both experiments. In addition, the instrument's sensitivity at dry conditions could be tested showing that the instrument sensitivity



drops by nearly an order of magnitude in the absence of water vapor. It is therefore beneficial to add water to the ion flow tube to maintain an high instrument sensitivity at very dry conditions of the sampled air.

The water vapor dependence of the sensitivity can be parameterized by a third order polynomial (Eq. 2) for water vapor mixing ratios higher than 0.15 %. This is typically sufficient for atmospheric conditions. At lower water vapor mixing ratios as experienced in the chamber experiments the parameterization in Eq. 3 provides a good fit. S is the signal normalized by the primary ion, a , b , c , d are the fit parameters and H_2O is the absolute water vapor mixing ratio.

$$S = a \times H_2O^3 + b \times H_2O^2 + c \times H_2O + d \quad : \quad H_2O \geq 0.15\% \quad (2)$$

$$S = c \times H_2O^{-0.4} + b \times H_2O + a \quad : \quad H_2O < 0.15\% \quad (3)$$

For the chamber experiments, the chamber air was humidified at the beginning of each experiments. At that time, no HO_2 is expected to be present in the chamber. Therefore, the increase in the background signal that has the same water vapor dependence as the sensitivity (see next section) can be used to determine the relative change of the sensitivity on water vapor on a daily basis. All HO_2 data from the chamber experiments shown in Sect. 3.4 were evaluated by applying this procedure.

During the series of chamber experiments presented in Sect. 3.4, calibrations were done in-between the experiments. In the middle of the series of experiments (6 June), settings of the instrument were tuned changing the sensitivity of the instrument. In total 6 calibrations were performed.

3.2 Precision of the HO_2 measurement

The precision of the instrument can be demonstrated by the Allan deviation plot shown in Fig. 4. 10 hours of measurement were used for this analysis while the instrument sampled from the calibration source that was operated at constant conditions. As mentioned above only the signal at mass-to-charge ratio 112 is used for simplicity. The calibration source constantly produced $2.5 \times 10^9 HO_2$ molecules cm^{-3} . A minimum integration time of 4 s was used for the evaluation, resulting in an Allan deviation of $1.7 \times 10^8 HO_2$ molecules cm^{-3} . With increasing integration time, the Allan deviation follows Gaussian noise demonstrating the statistically nature of the instrument's noise. An Allan deviation of $4.5 \times 10^7 HO_2$ molecules cm^{-3} is achieved, if the measurement is averaged over 60 s. This is a sufficient detection limit for atmospheric measurements. Lower detection limits can be achieved, if, for example, an integration time of 10 min is acceptable. In addition, the use of both isotopic signals at mass-to-charge ratio 112 and 114 would lower the detection limit by a factor of $\sqrt{2}$.

3.3 Instrumental background

The instrumental background was characterized in experiments where the inlet was overflowed with humidified synthetic air. This was done either using the radical source as a flow tube when the UV lamp was off or during experiments in SAPHIR, when only humidified synthetic air was present in the chamber. As shown in Fig. 5, the background signal changes similarly




Figure 4. Allan deviation plot derived from sampling a constant HO_2 concentration of $2.5 \times 10^9 \text{ HO}_2 \text{ molecules cm}^{-3}$ over 10 hours. The Allan deviation demonstrates the precision of measurements depending on the integration time. The red line indicates the behavior of the Allan deviation, if the noise is only limited by Gaussian noise.

with water vapor for both experimental conditions. The shape of the water vapor dependence is consistent with the assumption that a constant HO_2 concentration ($1.5 \pm 0.2 \times 10^8 \text{ molecules cm}^{-3}$) is internally produced in the instrument, which is detected according to the water vapor dependence of the instrument sensitivity discussed above. Therefore, the background can be subtracted from the measured HO_2 concentration after applying the water vapor dependent calibration factor. The value of the background needs to be regularly determined. For chamber experiments reported here, the background signal was measured in the clean dark chamber at the start of each experiment.

In turn, the change in the background signal with changing water vapor reflects the relative change in the instrument sensitivity. This is especially relevant for the experiments in the SAPHIR chamber, because the chamber air was humidified starting from dry synthetic air at the start of the experiments. Once the water addition was started the signal was rising steep and decreased slightly at higher water concentration, as shown in Fig. 5. No trend of the background signal over a period of 2 month was observed. The day-to-day variability of the background (in total 16 experiments) was within a range of $\pm 12\%$ during 2 months of measurements at the chamber.

3.3.1 Potential interference from ozone

Ozone is known to be an interference in some HO_2 LIF instruments due to the photolysis of O_3 by the 308 nm excitation laser (Holland et al., 2003). The potential ozone interference in the CIMS HO_2 detection was investigated in laboratory experiments. Ozone was added to humidified synthetic air (water vapor mixing ratios 0.2 and 2.6 %). For both conditions no increase of the background signal could be observed for ozone mixing ratios of up to 400 ppbv.




Figure 5. The background HO_2 measurement in the SAPHIR chamber derived during the humidification of the clean chamber and the background measured during the laboratory calibration supplying humidified synthetic air. The red line shows a third order polynomial fit representing Eq. 2, that can be used to corrected the instrument sensitivity at water mixing ratios higher than 0.15 %. For lower mixing ratios the orange fit is used representing Eq. 3.

During experiments in the SAPHIR chamber, instrument background effects can only be determined for periods of the experiments without the presence of reactants, when no HO_2 was present. Typically, ozone was added in a concentration of 100 to 200 ppbv. Although no artefacts were found in the laboratory characterization, an increase in the background upon ozone addition was observed in two of 12 experiments in SAPHIR. For both experiments, the chamber was first humidified and ozone was added afterwards. This appears as an increased intercept of 2.3×10^8 and 1.0×10^8 HO_2 molecules cm^{-3} in the linear regression between LIF and CIMS HO_2 data for the experiments of 21 June and 26 June (Fig. 7), respectively. The data of the LIF instrument were corrected for a maximum ozone interference of 0.05×10^8 and 0.15×10^8 HO_2 molecules cm^{-3} on these days, respectively. This correction is much smaller than the HO_2 concentration observed by the CIMS instrument, so that it can be excluded that differences are due to systematic errors in the data of the LIF instrument.

In the correlation plot (Fig. 8), including all experiments, this additional background was subtracted. The increased background due to the ozone addition has to be investigated in further chamber experiments. Because no direct connection between the occurrence of this interference and chemical conditions in the experiments is observed, it might be related to instrumental effects that could vary with time such as cleanness of the ion flow tube walls.




Figure 6. Time series plot for the HO_2 concentrations measured by the CIMS and the LIF instrument during the photo-oxidation experiment at 19 June 2017 in the SAPHIR chamber. The gray shaded area indicates that the chamber roof was closed. The vertical lines are showing the injection time of additional reactants, in case of water the injection took longer indicated by a broader line.

3.4 Comparison of CIMS and LIF HO_2 measurements

A time series for a typical experiment is shown in Fig. 6. The HO_2 production was initiated with the injection of ozone and the opening of the chamber roof providing UV light to the chamber. An addition of CO further boosted the HO_2 production, which dropped upon closing of the roof. However, HO_2 was still produced via radical chemistry in the dark. After the injection of water the CIMS shows a stable signal with a small offset. During the experiment the LIF and CIMS data reveal a good correlation having similar errors. This experiment was performed without the addition of a volatile organic compound (VOC), as well as, two other experiments marked with "None" in Fig. 7.

Figure 7 displays the correlation between HO_2 measurements by the CIMS and the LIF instrument for all day-long photo-oxidation experiment in the SAPHIR chamber performed in this study. The chemical composition was varied between experiments by changing for example the NO mixing ratio. The different chemical conditions during the experiments allows for checking for potential interferences. High NO concentrations of up to 3 ppbv were reached by injecting NO to the chamber air on 31.05 and 02.06, and up to 80 ppbv NO_2 was added on 23.06. The NO_2 interference test was performed injecting NO_2 in the dark, dry chamber. No further photo-chemistry experiments was done on this particular day. No systematic change in the relation between HO_2 data from the two instruments is observed in these cases (Fig. 7). In general, no interference from VOCs (Isoprene, ISOPOOH and reaction products) are observed, except for experiments with IEPOX injections. IEPOX was detected on m/z 197 as $\text{Br}^- \cdot \text{IEPOX}$ ion cluster, but the instrument was not calibrated for IEPOX. Nevertheless, this mass trace can be used to correct the HO_2 measurement for the interference from IEPOX. The HO_2 signal observed during the




Figure 7. Correlation between HO₂ measurements by the CIMS and LIF instruments for individual chamber experiments. Labels in the plots indicate the specific VOC injected into the chamber.

injection of IEPOX can be attributed to the interference from IEPOX, because IEPOX was injected in the dark chamber so that no HO₂ is expected to be present. This gives the relationship between the signal observed at the IEPOX mass (m/z 197) to the interference signal from IEPOX at the HO₂ mass (m/z 112). During the photo-oxidation of IEPOX, when also HO₂ is present, the interference signal can be subtracted from the signal at the HO₂ mass by scaling the initial interference signal by the relative change on m/z 197. The correction improves the correlation of the CIMS and the LIF but the absolute agreement is still not as good (slope of the regression 0.86; coefficient of determination 0.79) compared to the other experiments. However, a correction was performed for all experiments with IEPOX injection. The corrections are in the order of or smaller than the HO₂ measurements, and works best for the experiment with the lowest IEPOX concentration. It is worth noting that IEPOX concentrations were at least 10 times higher than typically found in the atmosphere. Kaiser et al. (2016) found IEPOX concentrations of 1 ppbv during a campaign in a forest in the South-East US where isoprene, the precursor of IEPOX, was the



dominant organic species. Therefore no significant interference for atmospheric measurements by the CIMS instrument are expected from IEPOX.

During experiments with ISOPOOH, HO₂ measurements by the LIF instrument showed higher values than HO₂ measured by the CIMS instrument (slope of linear regression of 0.78; coefficient of correlation $R^2 = 0.68$). Further experiments will be needed to investigate if ISOPOOH could cause an interference in the LIF instrument. Like in the case of IEPOX, ISOPOOH concentrations were much higher (several ppbv) than typically found in the atmosphere (less than 1 ppbv Kaiser et al. (2016)), so that no significant impact for atmospheric conditions is expected.

All concurrent measurements of the two instruments for HO₂ by CIMS and LIF, in the photo-oxidation experiments are summarized in the correlation plot shown in Fig. 8. In general, the correlation fit shows that there is an excellent agreement of both instruments giving a slope of linear regression of 1.07 and the linear correlation coefficient R^2 is 0.87. Experiments investigating the photo-oxidation of IEPOX and ISOPOOH are color-coded and are excluded from the correlation fit. However, using all data for the correlation fit leads to similar result (slope of linear regression of 1.05; coefficient of correlation $R^2 = 0.89$).

Correlation of individual experiments (Fig. 7, e.g. 21 June and 26 June) give partly significant offsets in the regression analysis of up to $2.3 \times 10^8 \text{ cm}^{-3} \text{ HO}_2$. One possible reason could be the procedure, how the water vapor dependence of the instrument sensitivity was derived. This was done by using the relative change of the a presumably constant instrumental HO₂ background during the humidification of the clean chamber air. The water vapor concentration was measured at a different location in the chamber. Therefore, there is the potential that the water vapor concentration measurement in the chamber was not representative for the water concentration in the ion flow tube of the instrument. In this case, the determination of the relative change of the instrument's sensitivity would fail and could results in an offset in the evaluation of data during the experiments. As seen in the experiment, shown in Fig. 6, there is a small offset that starts with the humidification. To avoid this effect in the future, a humidity sensor will be implemented at the ion flow tube.

4 Conclusion and Outlook

Chemical ionization was applied to measure atmospheric HO₂ concentrations using bromide ions as reagent. Laboratory characterization experiments and measurements in the atmospheric simulation chamber SAPHIR in Jülich were used to check the instruments applicability for atmospheric measurements. The performance of the CIMS instrument is comparable with measurements by a laser-induced fluorescence instrument. A water vapor dependence of the instrument sensitivity needs to be taken into account in the evaluation of data because the sensitivity of the instrument changes by roughly a factor of 2 for atmospheric water vapor concentrations between 0.2 and 1.4 %. Also a water vapor dependent background signal is observed. The change of the background signal with increasing water vapor, however, is explained by the water vapor dependence of the sensitivity. Therefore, the assumption is that the background consists of constant HO₂ production in the instrument. This background was stable within $\pm 12 \%$ during two months of measurements and no further trend was identified. The background signal and the instrument sensitivity needs to be quantified on a daily basis. No significant interference from trace gases




Figure 8. Correlation plot for the HO_2 concentrations measured by the CIMS and the LIF instrument of all photo-oxidation experiments in the SAPHIR chamber. A linear fit is applied to the subset of data excluding experiments with IEPOX and ISOPOOH.

NO , NO_2 , O_3 , CO , isoprene and ISOPOOH were found for atmospheric conditions. Only for non-atmospheric high IEPOX concentrations of several ppbv artificial signals were found that scaled with the IEPOX concentration. The HO_2 measurements correlate well with the LIF measurements. A slope of the linear regression of 1.07 was determined and a linear correlation coefficient (R^2) of 0.87 was found. With a detection limit of 4.5×10^7 molecules cm^{-3} for a 60 s measurement the instrument

5 is suitable to measure typical HO_2 concentrations in the atmosphere.

Further improvements of the instrument sensitivity might be expected, if wall contact of the sampled air including HO_2 is further minimized. This could be achieved by applying a sheath flow of pure nitrogen along the surface of the ion flow-tube.

Data availability. Data of the experiments in the SAPHIR chamber used in this work is available on the EUROCHAMP data homepage (<https://data.eurochamp.org/>, last access: June 2018).



Appendix A: Experimental parameters

day	H ₂ O	O ₃	VOC	VOC conc.	NO	NO ₂	CO
29 May	1.8 %	140 ppbv	cis-IEPOX	3.2 ppbv	0.06 ppbv	1.2 ppbv	0.03 ppmv
31 May	1.6 %	50 ppbv	isoprene	5 ppbv	3.0 ppbv	3.2 ppbv	0.05 ppmv
01 June	1.6 %	140 ppbv	trans-IEPOX	1.7 ppbv	0.05 ppbv	1.0 ppbv	0.03 ppmv
02 June	1.7 %	30 ppbv	cis-IEPOX	4.3 ppbv	3.0 ppbv	1.8 ppbv	0.06 ppmv
14 June	1.8 %	170 ppbv	1,2-ISOPROOH	2.4 ppbv	0.06 ppbv	1.3 ppbv	0.02 ppmv
15 June	2.1 %	190 ppbv	1,2-ISOPROOH	0.7 ppbv	0.04 ppbv	1.0 ppbv	0.03 ppmv
19 June	1.9 %	70 ppbv	none		0.14 ppbv	2.1 ppbv	0.9 ppmv
20 June	1.9 %	160 ppbv	1,2-ISOPROOH	1.9 ppbv	0.04 ppbv	1.2 ppbv	0.04 ppmv
21 June	0.03 %	170 ppbv	1,2-ISOPROOH, 4,3-ISOPROOH	6.5 ppbv 1 ppbv	0.01 ppbv	0.4 ppbv	0.15 ppmv
22 June	2.0 %	170 ppbv	4,3-ISOPROOH	9 ppbv	0.04 ppbv	1.0 ppbv	0.1 ppmv
24 June	1.4 %	160 ppbv	none		0.06 ppbv	0.8 ppbv	3.0 ppmv
26 June	1.6 %	170 ppbv	none		0.05 ppbv	0.9 ppbv	0.02 ppmv

Table A1. The maximum concentration of reactants present in the reaction chamber during the experiments used for the correlation plots show in 7. The VOC concentrations for ISOPROOH and IEPOX are preliminary data and are having an higher uncertainty.

Competing interests. The authors declare that they have no conflict of interest.

Acknowledgements. This project has received funding from the European Research Council (ERC) (SARLEP grant agreement No. 681529) and from the European Research Council(ERC) under the 15 European Union's Horizon 2020 research and innovation programme (Eurochamp
 5 2020 grant agreement No.730997). Thanks to Patrick Veres from NOAA for discussion and first ideas regarding the HO₂ CIMS detection. We like to acknowledge Jean Rivera-Ros from the Harvard University and David Reimer from our institute for the synthesis of the VOCs used for this study. Thanks to Martin Breitenlechner, Alexander Zaytsev and Frank N. Keutsch from School of Engineering and Applied Sciences and Department of Chemistry and Chemical Biology, Harvard University, Cambridge, MA, USA for the PTR measurements performed.



References

- Albrecht, S., Klopotoski, S., Derpmann, V., Klee, S., Brockmann, K. J., Stroh, F., and Benter, T.: Studies of the mechanism of the cluster formation in a thermally sampling API MS, *Rev. Sci. Instrum.*, 85, 014 102, <https://doi.org/10.1063/1.4854855>, 2014.
- Aloisio, S. and Francisco, J. S.: Existence of a Hydroperoxy and Water (HO₂-H₂O) Radical Complex, *J. Phys. Chem. A*, 102, 1899–1902, <https://doi.org/10.1021/jp972173p>, 1998.
- Andrés-Hernández, M., Stone, D., Brookes, D., Commane, R., Reeves, C., Huntrieser, H., Heard, D., Monks, P., Burrows, J., Schlager, H., et al.: Peroxy radical partitioning during the AMMA radical intercomparison exercise, *Atmos. Chem. Phys.*, 10, 10 621–10 638, <https://doi.org/10.5194/acpd-10-8447-2010>, 2010.
- Apel, E. C., Brauers, T., Koppmann, R., Bandowe, B., Boßmeyer, J., Holzke, C., Tillmann, R., Wahner, A., Wegener, R., Brunner, A., Jocher, M., Ruuskanen, T., Spirig, C., Steigner, D., Steinbrecher, R., Gomez Alvarez, E., Müller, K., Burrows, J. P., Schade, G., Solomon, S. J., Ladstätter-Weißenmayer, A., Simmonds, P., Young, D., Hopkins, J. R., Lewis, A. C., Legreid, G., Reimann, S., Hansel, A., Wisthaler, A., Blake, R. S., Ellis, A. M., Monks, P. S., and Wyche, K. P.: Intercomparison of oxygenated volatile organic compound measurements at the SAPHIR atmosphere simulation chamber, *J. Geophys. Res. Atmos.*, 113, <https://doi.org/10.1029/2008JD009865>, D20307, 2008.
- Burkert, J., Andrés-Hernández, M.-D., Stöbener, D., Burrows, J. P., Weissenmayer, M., and Kraus, A.: Peroxy radical and related trace gas measurements in the boundary layer above the Atlantic Ocean, *J. Geophys. Res. Atmos.*, 106, 5457–5477, <https://doi.org/10.1029/2000jd900613>, 2001.
- Caldwell, G., Masucci, J., and Ikonomou, M.: Negative ion chemical ionization mass spectrometry-binding of molecules to bromide and iodide anions, *J. Mass Spectrom.*, 24, 8–14, <https://doi.org/10.1002/oms.1210240103>, 1989.
- Cantrell, C. A., Stedman, D. H., and Wendel, G. J.: Measurement of atmospheric peroxy radicals by chemical amplification, *Anal. Chem.*, 56, 1496–1502, <https://doi.org/10.1021/ac00272a065>, 1984.
- Clemitshaw, K. C., Carpenter, L. J., Penkett, S. A., and Jenkin, M. E.: A calibrated peroxy radical chemical amplifier for ground-based tropospheric measurements, *J. Geophys. Res. Atmos.*, 102, 25 405–25 416, <https://doi.org/10.1029/97jd01902>, 1997.
- Derpmann, V., Albrecht, S., and Benter, T.: The role of ion-bound cluster formation in negative ion mass spectrometry, *Rapid Commun. Mass Spectrom.*, 26, 1923–1933, <https://doi.org/10.1002/rcm.6303>, 2012.
- Dorn, H.-P., Apodaca, R. L., Ball, S. M., Brauers, T., Brown, S. S., Crowley, J. N., Dubé, W. P., Fuchs, H., Häseler, R., Heitmann, U., Jones, R. L., Kiendler-Scharr, A., Labazan, I., Langridge, J. M., Meinen, J., Mentel, T. F., Platt, U., Pöhler, D., Rohrer, F., Ruth, A. A., Schlosser, E., Schuster, G., Shillings, A. J. L., Simpson, W. R., Thieser, J., Tillmann, R., Varma, R., Venables, D. S., and Wahner, A.: Intercomparison of NO₃ radical detection instruments in the atmosphere simulation chamber SAPHIR, *Atmos. Meas. Tech.*, 6, 1111–1140, <https://doi.org/10.5194/amt-6-1111-2013>, 2013.
- Edwards, G. D., Cantrell, C. A., Stephens, S., Hill, B., Goyea, O., Shetter, R. E., Mauldin, R. L., Kosciuch, E., Tanner, D. J., and Eisele, F. L.: Chemical ionization mass spectrometer instrument for the measurement of tropospheric HO₂ and RO₂, *Anal. Chem.*, 75, 5317–5327, <https://doi.org/10.1021/ac034402b>, 2003.
- Fuchs, H., Brauers, T., Dorn, H.-P., Harder, H., Häseler, R., Hofzumahaus, A., Holland, F., Kanaya, Y., Kajii, Y., Kubistin, D., Lou, S., Martinez, M., Miyamoto, K., Nishida, S., Rudolf, M., Schlosser, E., Wahner, A., Yoshino, A., and Schurath, U.: Technical Note: Formal blind intercomparison of HO₂ measurements in the atmosphere simulation chamber SAPHIR during the HOxComp campaign, *Atmos. Chem. Phys.*, 10, 12 233–12 250, <https://doi.org/10.5194/acp-10-12233-2010>, 2010.



- Fuchs, H., Bohn, B., Hofzumahaus, A., Holland, F., Lu, K., Nehr, S., Rohrer, F., and Wahner, A.: Detection of HO₂ by laser-induced fluorescence: calibration and interferences from RO₂ radicals, *Atmos. Meas. Tech.*, 4, 1209, <https://doi.org/10.5194/amt-4-1209-2011>, 2011.
- Fuchs, H., Dorn, H.-P., Bachner, M., Bohn, B., Brauers, T., Gomm, S., Hofzumahaus, A., Holland, F., Nehr, S., Rohrer, F., Tillmann, R.,
5 and Wahner, A.: Comparison of OH concentration measurements by DOAS and LIF during SAPHIR chamber experiments at high OH reactivity and low NO concentration, *Atmos. Meas. Tech.*, 5, 1611–1626, <https://doi.org/10.5194/amt-5-1611-2012>, 2012.
- Green, T. J., Reeves, C. E., Fleming, Z. L., Brough, N., Rickard, A. R., Bandy, B. J., Monks, P. S., and Penkett, S. A.: An improved dual channel PERCA instrument for atmospheric measurements of peroxy radicals, *J. Environ. Monitor.*, 8, 530–536, <https://doi.org/10.1039/b514630e>, 2006.
- 10 Hanke, M., Uecker, J., Reiner, T., and Arnold, F.: Atmospheric peroxy radicals: ROXMAS, a new mass-spectrometric methodology for speciated measurements of HO₂ and RO₂ and first results, *Int. J. Mass Spectrom.*, 213, 91–99, [https://doi.org/10.1016/S1387-3806\(01\)00548-6](https://doi.org/10.1016/S1387-3806(01)00548-6), 2002.
- Harrison, A. G.: Chemical ionization mass spectrometry, CRC press, Boca Raton, 2nd edn., <https://doi.org/10.1006/rwsp.2000.0357>, 1992.
- Hastie, D. R., Weissenmayer, M., Burrows, J. P., and Harris, G. W.: Calibrated chemical amplifier for atmospheric RO_x measurements, *Anal.*
15 *Chem.*, 63, 2048–2057, <https://doi.org/10.1021/ac00018a029>, 1991.
- Holland, F., Hofzumahaus, A., Schäfer, J., Kraus, A., and Pätz, H.-W.: Measurements of OH and HO₂ radical concentrations and photolysis frequencies during BERLIOZ, *J. Geophys. Res. Atmos.*, 108, <https://doi.org/10.1029/2001jd001393>, 2003.
- Hornbrook, R. S., Crawford, J. H., Edwards, G. D., Goyea, O., Mauldin III, R. L., Olson, J. S., and Cantrell, C. A.: Measurements of tropospheric HO₂ and RO₂ by oxygen dilution modulation and chemical ionization mass spectrometry, *Atmos. Meas. Tech.*, 4, 735–756,
20 <https://doi.org/10.5194/amt-4-735-2011>, 2011.
- Kaiser, J., Skog, K. M., Baumann, K., Bertman, S. B., Brown, S. B., Brune, W. H., Crounse, J. D., de Gouw, J. A., Edgerton, E. S., Feiner, P. A., Goldstein, A. H., Koss, A., Misztal, P. K., Nguyen, T. B., Olson, K. F., St. Clair, J. M., Teng, A. P., Toma, S., Wennberg, P. O., Wild, R. J., Zhang, L., and Keutsch, F. N.: Speciation of OH reactivity above the canopy of an isoprene-dominated forest, *Atmos. Chem. Phys.*, 16, 9349–9359, <https://doi.org/10.5194/acp-16-9349-2016>, 2016.
- 25 Kanno, N., Tonokura, K., Tezaki, A., and Koshi, M.: Water Dependence of the HO₂ Self Reaction: Kinetics of the HO₂-H₂O Complex, *J. Phys. Chem. A*, 109, 3153–3158, <https://doi.org/10.1029/2005jd006805>, 2005.
- Klee, S., Derpmann, V., Wißdorf, W., Klopotoski, S., Kersten, H., Brockmann, K. J., Benter, T., Albrecht, S., Bruins, A. P., Dousty, F., et al.: Are clusters important in understanding the mechanisms in atmospheric pressure ionization? Part 1: Reagent ion generation and chemical control of ion populations, *J. Am. Soc. Mass Spectrom.*, 25, 1310–1321, <https://doi.org/10.1007/s13361-014-0891-2>, 2014.
- 30 Lew, M. M., Dusanter, S., and Stevens, P. S.: Measurement of interferences associated with the detection of the hydroperoxy radical in the atmosphere using laser-induced fluorescence, *Atmos. Meas. Tech.*, 11, 95–108, <https://doi.org/10.5194/amt-11-95-2018>, 2018.
- Mihele, C. M. and Hastie, D. R.: Optimized Operation and Calibration Procedures for Radical Amplifier-Type Detectors, *J. Atmos. Ocean. Tech.*, 17, 788–794, [https://doi.org/10.1175/1520-0426\(2000\)017<0788:OOACPF>2.0.CO;2](https://doi.org/10.1175/1520-0426(2000)017<0788:OOACPF>2.0.CO;2), 2000.
- Reiner, T., Hanke, M., and Arnold, F.: Atmospheric peroxy radical measurements by ion molecule reaction mass spectrometry:
35 A novel analytical method using amplifying chemical conversion to sulfuric acid, *J. Geophys. Res. Atmos.*, 102, 1311–1326, <https://doi.org/10.1029/96JD02963>, 1997.
- Rohrer, F., Bohn, B., Brauers, T., Brüning, D., Johnen, F.-J., Wahner, A., and Kleffmann, J.: Characterisation of the photolytic HONO-source in the atmosphere simulation chamber SAPHIR, *Atmos. Chem. Phys.*, 5, 2189–2201, <https://doi.org/10.5194/acp-5-2189-2005>, 2005.



Sadanaga, Y., Matsumoto, J., Sakurai, K.-i., Isozaki, R., Kato, S., Nomaguchi, T., Bandow, H., and Kajii, Y.: Development of a measurement system of peroxy radicals using a chemical amplification/laser-induced fluorescence technique, *Rev. Sci. Instrum.*, 75, 864–872, <https://doi.org/10.1063/1.1666985>, 2004.

5 Sanchez, J., Tanner, D. J., Chen, D., Huey, L. G., and Ng, N. L.: A new technique for the direct detection of HO₂ radicals using bromide chemical ionization mass spectrometry (Br-CIMS): initial characterization, *Atmos. Meas. Tech.*, 9, 3851–3861, <https://doi.org/10.5194/amt-2016-117>, 2016.

Stone, D. and Rowley, D. M.: Kinetics of the gas phase HO₂ self-reaction: Effects of temperature, pressure, water and methanol vapours, *Phys. Chem. Chem. Phys.*, 7, 2156–2163, <https://doi.org/10.1039/b502673c>, 2005.

10 Tan, Z., Fuchs, H., Lu, K., Hofzumahaus, A., Bohn, B., Broch, S., Dong, H., Gomm, S., Häsel, R., He, L., Holland, F., Li, X., Liu, Y., Lu, S., Rohrer, F., Shao, M., Wang, B., Wang, M., Wu, Y., Zeng, L., Zhang, Y., Wahner, A., and Zhang, Y.: Radical chemistry at a rural site (Wangdu) in the North China Plain: observation and model calculations of OH, HO₂ and RO₂ radicals, *Atmos. Chem. Phys.*, 17, 663–690, <https://doi.org/10.5194/acp-17-663-2017>, 2017.

15 Veres, P., Roberts, J., Wild, R., Edwards, P., Brown, S., Bates, T., Quinn, P., Johnson, J., Zamora, R., and de Gouw, J.: Peroxynitric acid (HO₂NO₂) measurements during the UBWOS 2013 and 2014 studies using iodide ion chemical ionization mass spectrometry, *Atmos. Chem. Phys.*, 15, 8101–8114, <https://doi.org/10.5194/acpd-15-3629-2015>, 2015.

Whalley, L. K., Blitz, M. A., Desservettaz, M., Seakins, P. W., and Heard, D. E.: Reporting the sensitivity of laser-induced fluorescence instruments used for HO₂ detection to an interference from RO₂ radicals and introducing a novel approach that enables HO₂ and certain RO₂ types to be selectively measured, *Atmos. Meas. Tech.*, 6, 3425–3440, <https://doi.org/10.5194/amt-6-3425-2013>, 2013.